



检验报告



Page 1 of 3

Information Provided by Client	Applicant	ELOWEL INC. 460 kent avenue brooklyn n.y. 11249		Contact	---
	Manufacturer	----- -----		Tel.	---
	Information of Submitted Sample	Sample Name	Gown (Level 2)	Trademark	-----
		Sample Count	3Piece(s)		
		Size	-----	Colour	-----
Quality Grade		-----	Safety Category	-----	
	Style No. or Order No.	LFT201			
Test Part Description	1# Blue Isolation Gown				
Test Type	Commission Test	Date of Submission	2020-06-23	Date of Checking	2020-06-29
Test Standards	ANSI/AAMI PB70:2012 Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities				
Conclusion	Test results and compliance refer to next page(s).				
Non-standard Check Methods	-----				
Uncertainty of Result	-----				
Remarks	Weight: 42gsm				

Approver

方倩

Checker

张静

Editor

李楠





检验报告

CTTT-WT20017471 Page 2 of 3

Test Items	Description	Unit	Standard Requirement	Results	Conclusions	Test Method/Remarks
1#						
Water Resistance: Hydrostatic Pressure	Area B-Seam	cmH ₂ O	≥20	22	Pass	AATCC 127-2018
	Area C-Seam	cmH ₂ O	≥20	34		
	Area A-Fabric	cmH ₂ O	≥20	145		
	Area B-Fabric	cmH ₂ O	≥20	159		
	Area C-Fabric	cmH ₂ O	≥20	135		
Water Resistance: Impact penetration	Area B-Seam	g	≤1.0	0	Pass	AATCC 42-2017e
	Area C-Seam	g	≤1.0	0		
	Area A-Fabric	g	≤1.0	0		
	Area B-Fabric	g	≤1.0	0		
	Area C-Fabric	g	≤1.0	0		

Blank Part Below



170011263663



170011260226



(2017)国认监认字(090)号

检 验 报 告



中国认可
国际互认
检测
TESTING
CNAS L0608

CTTT-WT20017471

Page 3 of 3

Sample

